

## Calibration of a type N thermocouple at 1000 °C: Furnace temperature

In the N-type thermocouple calibration at 1000 °C the temperature of the furnace is determined by two R-type thermocouples placed the same distance above and below the thermocouple to be calibrated. The temperature at the location of the hot junction of the thermocouple to be calibrated is given as the mean of the temperatures of the hot junction of the two reference thermocouples.

### Model Equation:

$$t_x = t_s + \delta t_s + C_S \times (\delta V_{IS1} + \delta V_{IS2} + \delta V_R) - C_S / C_{S0} \times \delta t_{0S} + \delta t_D + \delta t_F$$

### List of Quantities:

Quantity	Unit	Definition
$t_x$	°C	temperature to be measured of the furnace
$t_s$	°C	temperature of the reference thermometer in terms of voltage with cold junction at 0 °C. The function is given in the calibration certificate.
$\delta t_s$	°C	correction of the temperature of the reference thermocouples due to their calibration
$C_S$	°C/ $\mu$ V	sensitivities of the reference thermocouples for voltage at the measuring temperature of 1000 °C
$\delta V_{IS1}$	$\mu$ V	voltage correction obtained from the calibration of the voltmeter
$\delta V_{IS2}$	$\mu$ V	voltage correction due to the limited resolution of the voltmeter
$\delta V_R$	$\mu$ V	voltage correction due to contact effects of the reversing switch
$C_{S0}$	°C/ $\mu$ V	sensitivities of the reference thermocouples for voltage at the reference temperature of 0 °C
$\delta t_{0S}$	°C	temperature correction due to the deviation of the reference temperature from 0 °C
$\delta t_D$	°C	change of the values of the reference thermometers since their last calibration due to drift
$\delta t_F$	°C	temperature correction due to non-uniformity of the temperature of the furnace

$t_s$ :

Type A

Method of observation: Direct

Number of observations: 1

No.	Observation
1	1000.5 °C

Arithmetic Mean: 1000.5000 °C

Pooled Standard Deviation: 0.10 °C

Pooled Degrees of Freedom: 9

Standard Uncertainty: 0.1000 °C

**MEASUREMENTS:** The indications of the voltmeter are precalculated to reduce the effects of temperature drift in the thermal source and parasitic thermal voltages (see EAL-R2-S1:S5.15). The readings are combined to one observation only, which is the temperature of the furnace at the location of the thermocouple to be calibrated. In order to evaluate the uncertainty of measurement associated with the observation, a series of ten measurements has been previously undertaken at the same temperature of operation. It gave a pooled estimate of standard deviation for the temperature of the furnace.

**$\delta t_S$ :** Type B normal distribution  
 Value: 0.0 °C  
 Expanded Uncertainty: 0.3 °C  
 Coverage Factor: 2

REFERENCE STANDARDS: The reference thermocouples are supplied with calibration certificates that relate the temperature at their hot junction with their cold junction at 0 °C to the voltage across their wires. The expanded uncertainty of measurement at 1000 °C is  $U=0,3$  °C (coverage factor  $k=2$ )

**$C_S$ :** Constant  
 Value: 0.077 °C/ $\mu$ V

VOLTAGE SENSITIVITIES: The voltage sensitivities of the thermocouples have been taken from reference tables

**$\delta V_{iS1}$ :** Type B normal distribution  
 Value: 0.0  $\mu$ V  
 Expanded Uncertainty: 2.0  $\mu$ V  
 Coverage Factor: 2.0

CALIBRATION OF THE VOLTMETER: The voltmeter has been calibrated. Corrections to the measured voltages are made to all results. The calibration certificate gives a constant expanded uncertainty of measurement for voltages smaller than 50 mV of  $U=2$   $\mu$ V (coverage factor  $k=2$ )

**$\delta V_{iS2}$ :** Type B rectangular distribution  
 Value: 0.0  $\mu$ V  
 Halfwidth of Limits: 0.5  $\mu$ V

RESOLUTION OF THE VOLTMETER: A 4½ digit microvoltmeter has been used in its 10 mV range resulting in resolution limits of  $\pm 0,5$   $\mu$ V at each indication.

**$\delta V_R$ :** Type B rectangular distribution  
 Value: 0.0  $\mu$ V  
 Halfwidth of Limits: 2.0  $\mu$ V

PARASITIC VOLTAGES: Residual parasitic offset voltages due to the switch contacts have been estimated to be zero within  $\pm 2$   $\mu$ V.

**$C_{S0}$ :** Constant  
 Value: 0.189 °C/ $\mu$ V

VOLTAGE SENSITIVITIES: The voltage sensitivities of the thermocouples have been taken from reference tables.

**$\delta t_{OS}$ :** Type B rectangular distribution  
 Value: 0.0 °C  
 Halfwidth of Limits: 0.1 °C

REFERENCE TEMPERATURE: The temperature of the reference point of each thermocouple is known to be 0 °C within  $\pm 0,1$  °C.

**$\delta t_D$ :** Type B rectangular distribution  
 Value: 0.0 °C  
 Halfwidth of Limits: 0.3 °C

DRIFT OF THE REFERENCE STANDARDS: From previous calibrations the drift of the reference standards are estimated to be zero within the limits  $\pm 0,3$  °C.

**$\delta t_F$ :** Type B rectangular distribution  
 Value: 0.0 °C  
 Halfwidth of Limits: 1 °C

TEMPERATURE GRADIENTS: The temperature gradients inside the furnace have been measured. At 1000 °C deviations from non-uniformity of temperature in the region of measurement are within  $\pm 1$  °C.

### Uncertainty Budgets:

$t_x$ : temperature to be measured of the furnace

Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index
$t_s$	1000.5000 °C	0.1000 °C	normal	1.0	0.10 °C	2.4 %
$\delta t_s$	0.0 °C	0.1500 °C	normal	1.0	0.15 °C	5.5 %
$C_s$	0.077 °C/ $\mu$ V					
$\delta V_{IS1}$	0.0 $\mu$ V	1.000 $\mu$ V	normal	0.077	0.077 °C	1.4 %
$\delta V_{IS2}$	0.0 $\mu$ V	0.2887 $\mu$ V	rectangular	0.077	0.022 °C	0.1 %
$\delta V_R$	0.0 $\mu$ V	1.155 $\mu$ V	rectangular	0.077	0.089 °C	1.9 %
$C_{S0}$	0.189 °C/ $\mu$ V					
$\delta t_{0S}$	0.0 °C	0.05774 °C	rectangular	-0.41	-0.024 °C	0.1 %
$\delta t_D$	0.0 °C	0.1732 °C	rectangular	1.0	0.17 °C	7.3 %
$\delta t_F$	0.0 °C	0.5774 °C	rectangular	1.0	0.58 °C	81.2 %
$t_x$	1000.5000 °C	0.6409 °C				

### Results:

Quantity	Value	Expanded Uncertainty	Coverage factor	Coverage
$t_x$	1000.5 °C	1.3 °C	2.00	95% (t-table 95.45%)