


Calibration of a coaxial step attenuator at a setting of 30 dB (incremental loss)

Author: EA

This Example is taken from EA 4/02. See EA 4/02 Section S7 for more details.

The measurement involves the calibration of a coaxial step attenuator at 10 GHz using an attenuation measuring system containing a calibrated step attenuator which acts as the attenuation reference. The method of measurement involves the determination of the attenuation between a matched source and matched load. In this case the unknown attenuator can be switched between a setting of 0 dB and 30 dB and it is this change (called incremental loss) that is determined in the calibration process. The attenuation measuring system has a digital readout and an analogue null detector which is used to indicate the balance condition.



Model Equation:

$$L_X = L_S + \delta L_S + \delta L_D + \delta L_M + \delta L_K + \delta L_{ib} - \delta L_{ia} + \delta L_{0b} - \delta L_{0a}$$

List of Quantities:

Quantity	Unit	Definition
L_X	dB	attenuation of the attenuator to be calibrated
L_S	dB	$=L_{ib} - L_{ia}$ attenuation difference of reference attenuator at 30 dB and 0 dB
δL_S	dB	correction obtained from the calibration of the reference attenuator
δL_D	dB	change of the attenuation of the reference attenuator since its last calibration due to drift
δL_M	dB	correction due to mismatch loss
δL_K	dB	correction for leakage signals between input and output of the attenuator to be calibrated due to imperfect isolation
δL_{ib}	dB	correction due to the limited resolution of the reference detector at 30 dB
δL_{ia}	dB	correction due to the limited resolution of the reference detector at 0 dB
δL_{0b}	dB	correction due to the limited resolution of the null detector at 30 dB
δL_{0a}	dB	correction due to the limited resolution of the null detector at 0 dB

L_S: Type A
 Method of observation: Direct
 Number of observations: 4

No.	Observation
1	30.033 dB
2	30.058 dB
3	30.018 dB
4	30.052 dB

Arithmetic Mean: 30.040250 dB
 Standard Deviation: 0.018 dB
 Standard Uncertainty: $9.132 \cdot 10^{-3}$ dB
 Degrees of Freedom: 3

MEASUREMENTS: Four observations are made of the incremental loss of the attenuator to be calibrated between settings of 0 dB and 30 dB. (see EAL-R2-S1:S7.11)

δL_S: Type B normal distribution
 Value: 0.003 dB
 Expanded Uncertainty: 0.005 dB
 Coverage Factor: 2

REFERENCE ATTENUATOR: The calibration certificate for the reference attenuator gives a value of attenuation for the 30,000 dB setting at 10 GHz of 30,003 dB with an associated expanded uncertainty of 0,005 dB (coverage factor k=2). The correction of +0,003 dB with the associated expanded uncertainty of 0,005 dB (coverage factor k=2) is considered to be valid for attenuation settings of the reference attenuator which differ not more $\pm 0,1$ dB from the calibrated setting of 30,000 dB.

δL_D: Type B rectangular distribution
 Value: 0.0 dB
 Halfwidth of Limits: 0.002 dB

DRIFT OF THE REFERENCE: The drift of the attenuation of the reference attenuator is estimated from its calibration history to be zero with limits $\pm 0,002$ dB.

δL_M: Type B U-shaped distribution
 Value: 0.0 dB
 Halfwidth of Limits: 0.0283 dB

MISMATCH LOSS: (see EAL-R2-S1:S7.6)

δL_K: Type B rectangular distribution
 Value: 0.0 dB
 Halfwidth of Limits: 0.003 dB

LEAKAGE CORRECTION: Leakage signals through the attenuator to be calibrated have been estimated from the measurements at 0 dB setting to be at least 100 dB below the measurement signal. The correction for leakage signals is estimated from these findings to be within $\pm 0,003$ dB at the 30 dB setting.

δL_{ib}: Type B rectangular distribution
 Value: 0.0 dB
 Halfwidth of Limits: 0.0005 dB

RESOLUTION OF THE REFERENCE ATTENUATOR SETTING: The digital readout of the reference attenuator has a resolution of 0,001 dB from which the correction for resolution is estimated to be within ± 0.0005 dB.

S7	Calibration of a coaxial step attenuator at a setting of 30 dB (incremental loss)																																																																																									
<p>δL_{ia}: Type B rectangular distribution Value: 0.0 dB Halfwidth of Limits: 0.0005 dB</p> <p>RESOLUTION OF THE REFERENCE ATTENUATOR SETTING: The digital readout of the reference attenuator has a resolution of 0,001 dB from which the correction for resolution is estimated to be within ± 0.0005 dB.</p> <p>δL_{ob}: Type B normal distribution Value: 0.0 dB Expanded Uncertainty: 0.002 dB Coverage Factor: 1.0</p> <p>RESOLUTION OF THE NULL DETECTOR: The detector resolution was determined from a previous evaluation to have a standard deviation of 0,002 dB at each reading with assumed normal probability distribution.</p> <p>δL_{oa}: Type B normal distribution Value: 0.0 dB Expanded Uncertainty: 0.002 dB Coverage Factor: 1</p> <p>RESOLUTION OF THE NULL DETECTOR: The detector resolution was determined from a previous evaluation to have a standard deviation of 0,002 dB at each reading with assumed normal probability distribution.</p> <p>Uncertainty Budgets:</p> <p>L_X: attenuation of the attenuator to becalibrated</p> <table border="1" data-bbox="226 1059 1482 1592"> <thead> <tr> <th>Quantity</th> <th>Value</th> <th>Standard Uncertainty</th> <th>Distribution</th> <th>Sensitivity Coefficient</th> <th>Uncertainty Contribution</th> <th>Index</th> </tr> </thead> <tbody> <tr> <td>L_S</td> <td>30.040250 dB</td> <td>$9.132 \cdot 10^{-3}$ dB</td> <td>normal</td> <td>1.0</td> <td>$9.1 \cdot 10^{-3}$ dB</td> <td>16.6 %</td> </tr> <tr> <td>δL_S</td> <td>$3.000 \cdot 10^{-3}$ dB</td> <td>$2.500 \cdot 10^{-3}$ dB</td> <td>normal</td> <td>1.0</td> <td>$2.5 \cdot 10^{-3}$ dB</td> <td>1.2 %</td> </tr> <tr> <td>δL_D</td> <td>0.0 dB</td> <td>$1.155 \cdot 10^{-3}$ dB</td> <td>rectangular</td> <td>1.0</td> <td>$1.2 \cdot 10^{-3}$ dB</td> <td>0.3 %</td> </tr> <tr> <td>δL_M</td> <td>0.0 dB</td> <td>0.02001 dB</td> <td>U-distr.</td> <td>1.0</td> <td>0.020 dB</td> <td>79.7 %</td> </tr> <tr> <td>δL_K</td> <td>0.0 dB</td> <td>$1.732 \cdot 10^{-3}$ dB</td> <td>rectangular</td> <td>1.0</td> <td>$1.7 \cdot 10^{-3}$ dB</td> <td>0.6 %</td> </tr> <tr> <td>δL_{ib}</td> <td>0.0 dB</td> <td>$288.7 \cdot 10^{-6}$ dB</td> <td>rectangular</td> <td>1.0</td> <td>$290 \cdot 10^{-6}$ dB</td> <td>0.0 %</td> </tr> <tr> <td>δL_{ia}</td> <td>0.0 dB</td> <td>$288.7 \cdot 10^{-6}$ dB</td> <td>rectangular</td> <td>-1.0</td> <td>$-290 \cdot 10^{-6}$ dB</td> <td>0.0 %</td> </tr> <tr> <td>δL_{ob}</td> <td>0.0 dB</td> <td>$2.000 \cdot 10^{-3}$ dB</td> <td>normal</td> <td>1.0</td> <td>$2.0 \cdot 10^{-3}$ dB</td> <td>0.8 %</td> </tr> <tr> <td>δL_{oa}</td> <td>0.0 dB</td> <td>$2.000 \cdot 10^{-3}$ dB</td> <td>normal</td> <td>-1.0</td> <td>$-2.0 \cdot 10^{-3}$ dB</td> <td>0.8 %</td> </tr> <tr> <td>L_X</td> <td>30.04325 dB</td> <td>0.02242 dB</td> <td></td> <td></td> <td></td> <td></td> </tr> </tbody> </table> <p>Results:</p> <table border="1" data-bbox="226 1680 1407 1803"> <thead> <tr> <th>Quantity</th> <th>Value</th> <th>Expanded Uncertainty</th> <th>Coverage factor</th> <th>Coverage</th> </tr> </thead> <tbody> <tr> <td>L_X</td> <td>30.043 dB</td> <td>0.045 dB</td> <td>2.00</td> <td>95% (t-table 95.45%)</td> </tr> </tbody> </table>			Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index	L_S	30.040250 dB	$9.132 \cdot 10^{-3}$ dB	normal	1.0	$9.1 \cdot 10^{-3}$ dB	16.6 %	δL_S	$3.000 \cdot 10^{-3}$ dB	$2.500 \cdot 10^{-3}$ dB	normal	1.0	$2.5 \cdot 10^{-3}$ dB	1.2 %	δL_D	0.0 dB	$1.155 \cdot 10^{-3}$ dB	rectangular	1.0	$1.2 \cdot 10^{-3}$ dB	0.3 %	δL_M	0.0 dB	0.02001 dB	U-distr.	1.0	0.020 dB	79.7 %	δL_K	0.0 dB	$1.732 \cdot 10^{-3}$ dB	rectangular	1.0	$1.7 \cdot 10^{-3}$ dB	0.6 %	δL_{ib}	0.0 dB	$288.7 \cdot 10^{-6}$ dB	rectangular	1.0	$290 \cdot 10^{-6}$ dB	0.0 %	δL_{ia}	0.0 dB	$288.7 \cdot 10^{-6}$ dB	rectangular	-1.0	$-290 \cdot 10^{-6}$ dB	0.0 %	δL_{ob}	0.0 dB	$2.000 \cdot 10^{-3}$ dB	normal	1.0	$2.0 \cdot 10^{-3}$ dB	0.8 %	δL_{oa}	0.0 dB	$2.000 \cdot 10^{-3}$ dB	normal	-1.0	$-2.0 \cdot 10^{-3}$ dB	0.8 %	L_X	30.04325 dB	0.02242 dB					Quantity	Value	Expanded Uncertainty	Coverage factor	Coverage	L_X	30.043 dB	0.045 dB	2.00	95% (t-table 95.45%)	
Quantity	Value	Standard Uncertainty	Distribution	Sensitivity Coefficient	Uncertainty Contribution	Index																																																																																				
L_S	30.040250 dB	$9.132 \cdot 10^{-3}$ dB	normal	1.0	$9.1 \cdot 10^{-3}$ dB	16.6 %																																																																																				
δL_S	$3.000 \cdot 10^{-3}$ dB	$2.500 \cdot 10^{-3}$ dB	normal	1.0	$2.5 \cdot 10^{-3}$ dB	1.2 %																																																																																				
δL_D	0.0 dB	$1.155 \cdot 10^{-3}$ dB	rectangular	1.0	$1.2 \cdot 10^{-3}$ dB	0.3 %																																																																																				
δL_M	0.0 dB	0.02001 dB	U-distr.	1.0	0.020 dB	79.7 %																																																																																				
δL_K	0.0 dB	$1.732 \cdot 10^{-3}$ dB	rectangular	1.0	$1.7 \cdot 10^{-3}$ dB	0.6 %																																																																																				
δL_{ib}	0.0 dB	$288.7 \cdot 10^{-6}$ dB	rectangular	1.0	$290 \cdot 10^{-6}$ dB	0.0 %																																																																																				
δL_{ia}	0.0 dB	$288.7 \cdot 10^{-6}$ dB	rectangular	-1.0	$-290 \cdot 10^{-6}$ dB	0.0 %																																																																																				
δL_{ob}	0.0 dB	$2.000 \cdot 10^{-3}$ dB	normal	1.0	$2.0 \cdot 10^{-3}$ dB	0.8 %																																																																																				
δL_{oa}	0.0 dB	$2.000 \cdot 10^{-3}$ dB	normal	-1.0	$-2.0 \cdot 10^{-3}$ dB	0.8 %																																																																																				
L_X	30.04325 dB	0.02242 dB																																																																																								
Quantity	Value	Expanded Uncertainty	Coverage factor	Coverage																																																																																						
L_X	30.043 dB	0.045 dB	2.00	95% (t-table 95.45%)																																																																																						
Date: 08/22/2007	File: S07_(picture).smu		Page 3 of 3																																																																																							